


## FAQ's About Analysis

1. **I get the following SICL32.dll message when I start the analysis software. Is there something wrong?**



Nothing's wrong. The problem is that Miz-18 is selected as the Tester in EddyAdmin. Simply select No Tester in EddyAdmin to make the message not appear anymore.